

KEYSIGHT
S/N: MY54301924
Cal Done: 24-Sep-2020
Due: 24-Sep-2020
Keysight Calibration


ES063A Network Analyzer

1 Active Ch/Trace 2 Response 3 Stimulus 4 Mkr/Analysis 5 Instr State

Scale 50 m/div

Tr1 S11 SWR 50.00 m/ Ref 1.000 [F2]

1	5.5889933 GHz	1.2398
>2	7.5863963 GHz	1.3156



Tr2 S21 Log Mag 10.00 dB/ Ref 0.000 dB [F2]

1	5.5889933 GHz	-0.2010 dB
>2	7.5863963 GHz	-0.2885 dB



Start 300 kHz

IFBW 70 kHz

Stop 8 GHz

Cor I

Meas

ExtRef

2024-08-22 15:01

Scale

Auto Scale

Auto Scale All

Divisions

10

Scale/Div

50 m/div

Reference Position

0 Div

Reference Value

1

Marker ->

Reference

Electrical Delay

Phase Offset

0 °

Return